

EC - DECLARATION OF CONFORMITY **96**

We,

Dwyer Instruments, Inc.
102 Indiana Highway 212
Michigan City, IN 46360 USA
+1-219-879-8868

declare under our sole responsibility that our Series 1900 Differential Pressure Switch to which this declaration relates is in conformity with the following Directives and harmonized standards:

Directive 2011/65/EU Restriction of Certain Hazardous Substances in Electrical and Electronic Equipment

Directive 2014/35/EU Low Voltage Directive (LVD)

IEC 61010-1:2010 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements

Model Nomenclature:

19XX-YY where XX may be 00, 10, 11, 20, or 21;
where YY may be 00, 0, 1, 5, 10, or 20.

The authorized representative located within the Community is:

Dwyer Instruments Ltd
Unit 16 The Wye Estate, London Road
High Wycombe, HP11 1LH-U.K.
+44 (0) 1494 461707

On behalf of Dwyer Instruments, Inc.,



Doug McCall
Authorized Signatory
Senior Regulatory Engineer
Title of Authorized Signatory
February 18, 2016
Date of Signature

Dwyer Instruments, Inc.
102 Indiana Highway 212
Michigan City, IN 46360 USA
+1-219-879-8868